

Herstelleraussage

Dokument Nr.:

14385898.02

Wir bestätigen hiermit für die Materialien und Gegenstände, die bei bestimmungsgemäßem Gebrauch in Kontakt mit Lebensmitteln kommen können, die Konformität mit den allgemeinen Anforderungen der

Verordnung (EG) Nr. 1935/2004 vom 27. Oktober 2004 über Materialien und Gegenstände, die dazu bestimmt sind, mit Lebensmitteln in Berührung zu kommen.

Dies gilt für folgende Typen

**Frontbündiger Druckmessumformer
S-11
in lebensmitteleuglicher Ausführung
mit Prozessanschluss G1B Hygienic**

gemäß dem gültigen Datenblatt,

PE 81.02

**Alle messstoffberührten Teile sind aus Edelstahl:
CrNi-Stahl 1.4435 (316L)**

Die eingesetzten metallischen Werkstoffe wurden getestet und halten die zulässigen Migrationsgrenzwerte ein.
(Ref.: Intertek SHAH00894786, 17-1271)

**Alle messstoffberührten Dichtungen sind aus
EPDM 70**


mit FDA, USP, 3A Zulassung.

Die Fertigungsprozesse der Geräte wurden nach „Guter Herstellungspraxis“, wie in der Verordnung (EG) Nr. 2023/2006 „über gute Herstellungspraxis für Materialien und Gegenstände, die dazu bestimmt sind, mit Lebensmitteln in Berührung zu kommen“ vom 22. Dezember 2006 beschrieben, durchgeführt.

Unterzeichnet für und im Namen von / *Signed for and on behalf of*

WIKA Alexander Wiegand SE & Co. KG

Klingenberg, 2020-07-22



Fokko Stuke,
Director Operations
Electronic Products – Industrial Instrumentation

Manufacturer's Statement

Document No.:

14385898.02

For the materials and articles which may come into contact with food when used as intended, we hereby confirm the conformity with the general requirements of

Regulation (EC) No 1935/2004 of 27 October 2004 on materials and articles intended to come into contact with food.

This applies to the following components:

**Flush pressure transmitter
S-11
Suitable for food with
with process connection G1B Hygienic**

according to the current data sheet

PE 81.02


**All wetted parts are made of stainless steel:
CrNi-steel 1.4435 (316L)**

The metallic materials used were tested and comply to the permissible migration limits.
(Ref.: Intertek SHAH00894786, 17-1271)

**All wetted sealing parts are made of
EPDM 70**

listed acc. to FDA, USP, 3A standards.

The manufacturing process has followed „Good Manufacturing Practice (GMP)“ as specified in the regulation (EC) No. 2023/2006“ on good manufacturing practice for materials and articles intended to come into contact with food“ of 22. December 2006.



Steffen Schlesiona,
Director Quality Management
Industrial Instrumentation